

For Export Only – Not For Sale in USA

ARK™ Pregabalin Urine Control

This ARK Diagnostics, Inc. package insert for the ARK Pregabalin Urine Control must be read carefully prior to use. Package insert instructions must be followed accordingly. Reliability of the assay results cannot be guaranteed if there are any deviations from the instructions in this package insert.

CUSTOMER SERVICE

 **ARK Diagnostics, Inc.**

48089 Fremont Blvd

Fremont, CA 94538 USA

Tel: 1-877-869-2320

Fax: 1-510-270-6298

customersupport@ark-tdm.com

www.ark-tdm.com



Emergo Europe

Molenstraat 15











2513 BH The Hague

The Netherlands

Tel: (31) (0) 70 345-8570

Fax: (31) (0) 70 346-7299

KEY TO SYMBOLS USED

	Batch code	 YYYY-MM-DD	Use by/Expiration date
	Catalog Number		Manufacturer
	Authorized Representative		CE Mark
	Consult Instructions for Use		Quality Control
	Temperature limitation		In Vitro Diagnostic Medical Device

1 NAME

ARK™ Pregabalin Urine Control

2 INTENDED USE

ARK Pregabalin Urine Control is intended for use in quality control of the ARK Pregabalin Urine Assay.

3 CONTENT

ARK Pregabalin Urine Control is comprised of a non-sterile, processed human urine matrix with the following concentrations of pregabalin:

REF	Product Description	Quality Control
5035-0003-00	ARK Pregabalin Control Pregabalin, human urine, stabilizer and sodium azide	Dropper vials
	LOW / Negative (250 ng/mL)	3 X 4 mL
	HIGH / Positive (750 ng/mL)	3 X 4 mL

Each laboratory should establish its own ranges for each new lot of controls. The ARK Pregabalin Urine Control is a non-assayed control intended for quality control of the ARK Pregabalin Urine Assay when run in either the qualitative or semi-quantitative mode.

In Qualitative Mode the Low Control should be Negative and the High Control should be Positive relative to the 500 ng/mL Cutoff Calibrator.

Controls are made with non-sterile, processed human urine free of pregabalin. Donors were non-reactive in tests for HIV 1/2, HBsAg, HCV, HIV-1 (NAT), HCV (NAT) and RPR.

4 WARNINGS AND PRECAUTIONS

- **For In Vitro Diagnostic Use. Harmful if swallowed.**
- Contains human urine. Handle as potentially infectious.
- Do not mix controls from different lot numbers.
- Use each lot as a set.
- Product contains ≤0.09% sodium azide. As a precaution, affected plumbing including instrumentation should be flushed adequately with water to mitigate the potential accumulation of explosive metal azides.

5 INSTRUCTIONS FOR USE

- For a complete summary and explanation of the Pregabalin Assay, refer to the package insert for the ARK Pregabalin Urine Assay.
- Controls are ready to use. Mix each level by gentle inversion before dispensing.
- Squeeze sufficient volume (~40µL/drop) into individual sample cups for each level. Consult instrument-specific sample volume requirements. Return caps to their original containers and keep tight.
- Store at 2-8°C. Use prior to the expiration date.

6 LIMITATIONS OF PROCEDURE

Accurate and reproducible results are dependent upon properly functioning instruments, reagents, calibrators, controls, storage of product as directed, and good laboratory technique.

7 TRADEMARKS

ARK™ is a trademark of ARK Diagnostics, Inc.

Other brand or product names are trademarks of their respective holders.