

For Criminal Justice and Forensic Use Only

ARK™ Methylphenidate Metabolite Calibrator

This ARK Diagnostics, Inc. package insert for the ARK Methylphenidate Metabolite Calibrator must be read carefully prior to use. Package insert instructions must be followed accordingly. Reliability of the assay results cannot be guaranteed if there are any deviations from the instructions in this package insert.

CUSTOMER SERVICE

 **ARK Diagnostics, Inc.**

48089 Fremont Blvd

Fremont, CA 94538 USA








Tel: 1-877-869-2320

Fax: 1-510-270-6298

customersupport@ark-tdm.com

www.ark-tdm.com

KEY TO SYMBOLS USED

	Batch code	 YYYY-MM-DD	Use by/Expiration date
	Catalog Number		Manufacturer
	Consult Instructions for Use		Calibrator
	Temperature limitation		

1 NAME

ARK™ Methylphenidate Metabolite Calibrator

2 INTENDED USE

This product is intended for Criminal Justice and Forensic Use Only.

The ARK Methylphenidate Metabolite Calibrator is intended for use in calibration of the ARK Methylphenidate Metabolite Assay.

3 CONTENT

The ARK Methylphenidate Metabolite Calibrator is composed of a non-sterile, processed human urine matrix with the following concentrations of Methylphenidate Metabolite. Negative and Cutoff calibrators may be obtained separately for qualitative analysis.

REF	Product Description	Quantity/Volume
5042-0005-00	ARK Methylphenidate Metabolite Calibrator Methylphenidate Metabolite, human urine, stabilizer and sodium azide	Dropper Vials
	A	0 ng/mL
	B	100 ng/mL
	C	200 ng/mL
	D	500 ng/mL
	E	1000 ng/mL

REF	Product Description	Quantity/Volume
5042-0005-01	ARK Methylphenidate Metabolite Calibrator A (Negative) Human urine, stabilizer and sodium azide	Dropper Vials
	Negative	0 ng/mL

REF	Product Description	Quantity/Volume
5042-0005-02	ARK Methylphenidate Metabolite Calibrator B (Cutoff) Methylphenidate Metabolite, human urine, stabilizer and sodium azide	Dropper Vials
	Cutoff	100 ng/mL

4 STANDARDIZATION

There is no internationally recognized standard for Methylphenidate Metabolite. A certified solution of Methylphenidate Metabolite is traceable to HPLC. ARK Methylphenidate Metabolite Calibrators are prepared by volumetric dilution of high purity Methylphenidate Metabolite into non-sterile, processed human urine free of Methylphenidate Metabolite.

Calibrators are made with non-sterile, processed human urine free of Methylphenidate Metabolite. Donors were non-reactive in tests for HIV 1/2, HBsAg, HCV, HIV-1 (NAT), HCV (NAT) and RPR.

5 WARNINGS AND PRECAUTIONS

- Not For *In Vitro* Diagnostic Use.
- Harmful if swallowed.
- Contains human urine. Handle as potentially infectious.
- Do not mix calibrators from different lot numbers.
- Use each lot as a set.
- Product contains ≤0.09% sodium azide. As a precaution, affected plumbing including instrumentation should be flushed adequately with water to mitigate the potential accumulation of explosive metal azides.

6 INSTRUCTIONS FOR USE

- For a complete summary and explanation of the Methylphenidate Metabolite Assay, refer to the package insert for the ARK Methylphenidate Metabolite Assay.
- Calibrators are ready to use. Mix each level by gentle inversion before dispensing.
- Squeeze sufficient volume (~40µL/drop) into individual sample cups for each level. Consult instrument-specific sample volume requirements. Return caps to their original containers and keep tight.
- Store at 2-8°C. Use prior to the expiration date.

7 PROCEDURE

Qualitative Results

Use the 100 ng/mL Calibrator B as a Cutoff Calibrator to distinguish negative and positive samples. Run the ARK Methylphenidate Metabolite Low (50 ng/mL) and High (150 ng/mL) Controls as Negative and Positive respectively. Report test results less than the response value for the Cutoff Calibrator as Negative. Report results equal to or greater than the response value for the Cutoff Calibrator as Positive.

Semiquantitative Results

Perform a 5-point calibration procedure; test calibrators in duplicate. Verify the calibration curve with the ARK Methylphenidate Metabolite Low (50 ng/mL) and High (150 ng/mL) quality controls according to the established laboratory quality assurance plan. Specimens with sample results above the highest ARK Methylphenidate Metabolite calibrator level (1000 ng/mL) may be diluted in ARK Methylphenidate Metabolite Calibrator A (Negative urine) and retested.

When to Re-Calibrate

- Whenever a new lot number of reagents is used
- Whenever indicated by quality control results
- Whenever required by standard laboratory protocols
- A stored calibration curve was effective up to at least 8 days based on supporting data

8 LIMITATIONS OF PROCEDURE

Accurate and reproducible results are dependent upon properly functioning instruments, reagents, calibrators, controls, storage of product as directed, and good laboratory technique.

9 TRADEMARKS

ARK[™] is a trademark of ARK Diagnostics, Inc.

Other brand or product names are trademarks of their respective holders.