

ARK Methotrexate II Dilution Buffer

This ARK Diagnostics, Inc. package insert for the ARK Methotrexate II Dilution Buffer must be read carefully prior to use. Package insert instructions must be followed accordingly. Reliability of the assay results cannot be guaranteed if there are any deviations from the instructions in this package insert.

Report any serious incident that has occurred in relation to the device to the manufacturer and the appropriate competent authority as applicable.

CUSTOMER SERVICE

 ARK Diagnostics, Inc.

48089 Fremont Blvd

Fremont, CA 94538 USA









Tel: 1-877-869-2320

Fax: 1-510-270-6298

customersupport@ark-tdm.com

www.ark-tdm.com

KEY TO SYMBOLS USED

	Batch code	 YYYY-MM-DD	Use by/Expiration date
	Catalog Number		Manufacturer
	<i>In Vitro</i> Diagnostic Medical Device		Temperature limitation
	Consult Instructions for Use		Quality Control
Rx Only	For Prescription Use Only		


1 NAME

ARK Methotrexate II Dilution Buffer

2 INTENDED USE

ARK Methotrexate II Dilution Buffer is intended for dilution of specimens containing high concentrations of methotrexate for the ARK Methotrexate II Assay.

3 CONTENT

ARK Methotrexate II Dilution Buffer is comprised of a synthetic protein matrix. Its composition is equivalent to Calibrator A (zero), ARK Methotrexate II Calibrator,  5071-0002-00.

REF	Product Description	Quantity/Volume
5071-0004-00	ARK Methotrexate II Dilution Buffer Buffer, bovine serum albumin, and sodium azide	1 X 25 mL

4 WARNINGS AND PRECAUTIONS

- For *In Vitro* Diagnostic Use.
- Dilution Buffer contains ≤0.09% sodium azide.

5 INSTRUCTIONS FOR USE

- For a complete summary and explanation of the ARK Methotrexate II Assay, refer to the package insert for the ARK Methotrexate II Assay.
- Dilution Buffer is ready to use. Mix by gentle inversion before dispensing.
- Store at 2-8°C. Use prior to the expiration date.

6 PROCEDURE – MANUAL DILUTION PROTOCOL

The measurement range of the ARK Methotrexate II calibration curve is 0.030 - 1.300 µmol/L. Specimens and controls containing methotrexate in higher concentrations (>1.300 µmol/L), are assayed by dilution of the specimen and controls into the measurement range.

Manually dilute the high specimen or high range control with ARK Methotrexate II Dilution Buffer by preparing the appropriate ten-fold serial dilution as shown below.

Sample Volume	Dilution Matrix Volume	Dilution	Dilution Factor
50 µL	Undiluted sample	450 µL	1:10
50 µL	1:10 sample	450 µL	1:100
50 µL	1:100 sample	450 µL	1:1000

$$\text{Manual Dilution Factor} = \frac{\text{Volume of Specimen} + \text{Volume of Dilution Buffer}}{\text{Specimen Volume}}$$

Multiply the assayed result by the dilution factor. To convert µmol/L to µg/mL, divide the value obtained by the conversion factor of 2.2005.

7 TRADEMARKS

ARK™ is a trademark of ARK Diagnostics, Inc.

Other brand or product names are trademarks of their respective holders.