

ARK™ Lamotrigine Control

This ARK Diagnostics, Inc. package insert for the ARK Lamotrigine Control must be read carefully prior to use. Package insert instructions must be followed accordingly. Reliability of the assay results cannot be guaranteed if there are any deviations from the instructions in this package insert.











CUSTOMER SERVICE

 ARK Diagnostics, Inc.
48089 Fremont Blvd
Fremont, CA 94538 USA
Tel: 1-877-869-2320
Fax: 1-510-270-6298
customersupport@ark-tdm.com
www.ark-tdm.com



Ergo Europe
Prinsessegracht 20
2514 AP The Hague
The Netherlands


KEY TO SYMBOLS USED

	Batch code	 YYYY-MM-DD	Use by/Expiration date
	Catalog Number		Manufacturer
	Authorized Representative		CE Mark
	In Vitro Diagnostic Medical Device		Temperature limitation
	Consult Instructions for Use		Quality Control
Rx Only	For Prescription Use Only		

1 NAME

ARK™ Lamotrigine Control

2 INTENDED USE

ARK Lamotrigine Control is intended for use in quality control of the ARK Lamotrigine Assay,  5023-0001-00.

3 CONTENT

ARK Lamotrigine Control is comprised of a synthetic protein matrix with the following concentrations of lamotrigine:

REF	Product Description	Quantity/Volume
5023-0003-00	ARK Lamotrigine Control* Lamotrigine, buffer, bovine serum albumin, and preservatives (nominal level)	Dropper vials
	LOW (2.00 µg/mL)	1 X 4 mL
	MID (12.00 µg/mL)	1 X 4 mL
	HIGH (25.00 µg/mL)	1 X 4 mL

* To convert results from µg/mL lamotrigine to µmol/L lamotrigine, multiply µg/mL by 3.90. Lamotrigine levels become 7.80, 46.80, and 97.50 µmol/L for trilevel controls respectively.

Each laboratory should establish its own ranges for each new lot of controls.

4 WARNINGS AND PRECAUTIONS

- For **In Vitro Diagnostic** Use.
- Do not mix controls from different lot numbers.
- Use each lot as a set.

5 INSTRUCTIONS FOR USE

- For a complete summary and explanation of the lamotrigine assay, refer to the package insert for the ARK Lamotrigine Assay.
- Controls are ready to use. Mix each level by gentle inversion before dispensing.
- Squeeze sufficient volume (~40µL/drop) into individual sample cups for each level. Consult instrument-specific sample volume requirements. Return caps to their original containers and keep tight.
- Store vials at 2-8°C. Use prior to the expiration date. Once opened, vials may be stored at 2-8°C for 12 months within the expiration date.

6 LIMITATIONS OF PROCEDURE

Accurate and reproducible results are dependent upon properly functioning instruments, reagents, calibrators, controls, storage of product as directed, and good laboratory technique.

7 TRADEMARKS

ARK™ is a trademark of ARK Diagnostics, Inc.
Other brand or product names are trademarks of their respective holders.