

ARK™ Levetiracetam II Control

This ARK Diagnostics, Inc. package insert for the ARK Levetiracetam II Control must be read carefully prior to use. Package insert instructions must be followed accordingly. Reliability of the assay results cannot be guaranteed if there are any deviations from the instructions in this package insert.

CUSTOMER SERVICE

 ARK Diagnostics, Inc.











48089 Fremont Blvd
Fremont, CA 94538 USA
Tel: 1-877-869-2320
Fax: 1-510-270-6298

customersupport@ark-tdm.com
www.ark-tdm.com



Emergo Europe
Prinsessegracht 20
2514 AP The Hague
The Netherlands


KEY TO SYMBOLS USED

	Batch code	 YYYY-MM-DD	Use by/Expiration date
	Catalog Number		Manufacturer
	Authorized Representative		CE Mark
	Consult Instructions for Use		Quality Control
	Temperature limitation		<i>In Vitro</i> Diagnostic Medical Device
Rx Only	For Prescription Use Only		

1 NAME

ARK™ Levetiracetam II Control

2 INTENDED USE

ARK Levetiracetam II Control is intended for use in quality control of the ARK Levetiracetam II Assay,  5070-0001-00.

3 CONTENT

ARK Levetiracetam II Control is comprised of a synthetic protein matrix with the following concentrations of levetiracetam:

REF	Product Description	Quantity/Volume
5070-0003-00	ARK Levetiracetam II Control Levetiracetam, buffer, bovine serum albumin, and preservatives (nominal level)	Dropper Vials
	LOW (7.5 µg/mL)	1 X 4 mL
	MID (30.0 µg/mL)	1 X 4 mL
	HIGH (75.0 µg/mL)	1 X 4 mL

Each laboratory should establish its own ranges for each new lot of controls.

4 WARNINGS AND PRECAUTIONS

- For *In Vitro* Diagnostic Use.
- Do not mix controls from different lot numbers.
- Use each lot as a set.
- Controls contain ≤0.09% sodium azide.

5 INSTRUCTIONS FOR USE

- For a complete summary and explanation of the levetiracetam assay, refer to the package insert for the ARK Levetiracetam II Assay.
- Controls are ready to use. Mix each level by gentle inversion before dispensing.
- Squeeze sufficient volume (~40µL/drop) into individual sample cups for each level. Consult instrument-specific sample volume requirements. Return caps to their original containers and keep tight.
- Store at 2-8°C. Use prior to the expiration date.

6 LIMITATIONS OF PROCEDURE

Accurate and reproducible results are dependent upon properly functioning instruments, reagents, calibrators, controls, storage of product as directed, and good laboratory technique.

7 TRADEMARKS

ARK™ is a trademark of ARK Diagnostics, Inc.

Other brand or product names are trademarks of their respective holders.