

## ARK™ *Levetiracetam Control*

This ARK Diagnostics, Inc. package insert for the ARK Levetiracetam Control must be read carefully prior to use. Package insert instructions must be followed accordingly. Reliability of the assay results cannot be guaranteed if there are any deviations from the instructions in this package insert.

Report any serious incident that has occurred in relation to the device to the manufacturer and the appropriate competent authority as applicable. A Summary of Safety and Performance is available through Eudamed (European database on medical devices), SRN: US-MF-000023925.

### Customer Service




48089 Fremont Blvd  
 Fremont, CA 94538 USA  
 Tel: 1-877-869-2320  
 Fax: 1-510-270-6298  
 customersupport@ark-tdm.com  
 www.ark-tdm.com









**2797**





Emergo Europe  
 Westervoortsedijk 60  
 6827 AT Arnhem  
 The Netherlands



MedEnvoy Switzerland  
 Gotthardstrasse 28  
 6302 Zug  
 Switzerland

### Key to Symbols Used

	Batch code	 YYYY-MM-DD	Use by/Expiration date
	Catalog Number		Manufacturer
	Authorized Representative		CE Mark with notified body number

	In Vitro Diagnostic Medical Device		Temperature limitation
	Consult Instructions for Use		Quality Control
<b>Rx Only</b>	For Prescription Use Only		

© 2025, **ARK** Diagnostics, Inc.

Control Kit  **5024-0003-00**

## 1 Name

ARK<sup>TM</sup> *Levetiracetam Control*

## 2 Intended Use

ARK Levetiracetam Control is intended for use in quality control of the ARK Levetiracetam Assay, **REF** 5024-0001-00.

## 3 Content

ARK Levetiracetam Control is comprised of a synthetic protein matrix with the following concentrations of levetiracetam:

REF	Product Description	Quantity/Volume
5024-0003-00	<b>ARK Levetiracetam Control</b> Levetiracetam, buffer, bovine serum albumin, and preservatives (nominal level)	Dropper vials
	<b>LOW</b> (7.5 µg/mL)	1 X 4 mL
	<b>MID</b> (30.0 µg/mL)	1 X 4 mL
	<b>HIGH</b> (75.0 µg/mL)	1 X 4 mL

Each laboratory should establish its own ranges for each new lot of controls.

## 4 Warnings and Precautions

- For *In Vitro Diagnostic* Use.
- Do not mix controls from different lot numbers.
- Use each lot as a set.

## 5 Instructions For Use

- For a complete summary and explanation of the levetiracetam assay, refer to the package insert for the ARK Levetiracetam Assay.
- Controls are ready to use. Mix each level by gentle inversion before dispensing.
- Squeeze sufficient volume (~40µL/drop) into individual sample cups for each level. Consult instrument-specific sample volume requirements. Return caps to their original containers and keep tight.
- Store at 2-8°C. Once opened, use within 12 months and prior to the expiration date.

## 6 Limitations of Procedure

Accurate and reproducible results are dependent upon properly functioning instruments, reagents, calibrators, controls, storage of product as directed, and good laboratory technique.

## 7 Trademarks

ARK<sup>TM</sup> is a trademark of ARK Diagnostics, Inc.

Other brand or product names are trademarks of their respective holders.



**ARK Diagnostics, Inc.**  
**Fremont, CA 94538 USA**

Revised January 2025  
1600-0171-00 Rev 06