

For Export Only – Not For Sale in USA

ARK™ AB-PINACA Control

This ARK Diagnostics, Inc. package insert for the ARK AB-PINACA Control must be read carefully prior to use. Package insert instructions must be followed accordingly. Reliability of the assay results cannot be guaranteed if there are any deviations from the instructions in this package insert.

CUSTOMER SERVICE

 **ARK Diagnostics, Inc.**











48089 Fremont Blvd
Fremont, CA 94538 USA
Tel: 1-877-869-2320
Fax: 1-510-270-6298

customersupport@ark-tdm.com
www.ark-tdm.com



Emergo Europe
Prinsessegracht 20
2514 AP The Hague
The Netherlands

KEY TO SYMBOLS USED

	Batch code	 YYYY-MM-DD	Use by/Expiration date
	Catalog Number		Manufacturer
	Authorized Representative		CE Mark
	Consult Instructions for Use		Quality Control
	Temperature limitation		<i>In Vitro</i> Diagnostic Medical Device
Rx Only	For Prescription Use Only		

1 NAME

ARK™ AB-PINACA Control

2 INTENDED USE

The ARK AB-PINACA Control is intended for use in quality control of the ARK AB-PINACA Assay.

3 CONTENT

The ARK AB-PINACA Control is composed of a non-sterile, processed human urine matrix with the following target concentrations of AB-PINACA Pentanoic Acid.

REF	Product Description	Quantity/Volume
5055-0003-00	ARK AB-PINACA Control AB-PINACA Pentanoic Acid, human urine, stabilizer and sodium azide	Dropper Vials
	LOW / Negative (2.5 ng/mL)	2 X 10 mL
	HIGH / Positive (7.5 ng/mL)	2 X 10 mL

Traceability and Value Verification: A certified solution of AB-PINACA Pentanoic Acid is traceable to HPLC. Testing is performed with the ARK AB-PINACA Assay calibrated with the ARK AB-PINACA Calibrator.

Each laboratory should establish its own ranges for each new lot of controls based on its own test system and criteria.

The Low Control should be Negative and the High Control should be Positive relative to the 5 ng/mL Cutoff Calibrator.

Controls are made with non-sterile, processed human urine. Donors were non-reactive in tests for HIV 1/2, HBsAg, HCV, HIV-1 (NAT), HCV (NAT) and RPR.

4 WARNINGS AND PRECAUTIONS

- For *In Vitro* Diagnostic Use. For prescription use only.
- Harmful if swallowed.
- Contains human urine. Handle as potentially infectious.
- Do not mix controls from different lot numbers.
- Use each lot as a set.
- Product contains ≤0.09% sodium azide. As a precaution, affected plumbing including instrumentation should be flushed adequately with water to mitigate the potential accumulation of explosive metal azides.

5 INSTRUCTIONS FOR USE

- For a complete summary and explanation of the AB-PINACA Assay, refer to the package insert for the ARK AB-PINACA Assay.
- Controls are ready to use. Mix each level by gentle inversion before dispensing.
- Squeeze sufficient volume (~40µL/drop) into individual sample cups for each level. Consult instrument-specific sample volume requirements. Return caps to their original containers and keep tight.
- Store at 2-8°C. Use prior to the expiration date.

6 LIMITATIONS OF PROCEDURE

Accurate and reproducible results are dependent upon properly functioning instruments, reagents, calibrators, controls, storage of product as directed, and good laboratory technique.

7 TRADEMARKS

ARK™ is a trademark of ARK Diagnostics, Inc.

Other brand or product names are trademarks of their respective holders.