ARK Diagnostics, Inc.

ARK™ *SDMA Control*

This ARK Diagnostics, Inc. package insert for the ARK SDMA Control must be read carefully prior to use. Package insert instructions must be followed accordingly. Reliability of the assay results cannot be guaranteed if there are any deviations from the instructions in this package insert.

Customer Service

ARK Diagnostics, Inc.
48089 Fremont Blvd
Fremont, CA 94538 USA
Tel: 1-877-869-2320
Fax: 1-510-270-6298
customersupport@ark-tdm.com
www.ark-tdm.com

Key to Symbols Used

LOT	Batch code	YYYY-MM-DD	Use by/Expiration date
REF	Catalog Number		Manufacturer
	Consult Instructions for Use		Temperature limitation
CONTROL	Quality Control		

© 2026, ARK Diagnostics, Inc.

Control Kit REF 5085-0003-00

1 Name

ARK *SDMA Control*

2 Intended Use

ARK SDMA Control is intended for use in quality control of the ARK SDMA Assay.

3 Content

ARK SDMA Control is comprised of a synthetic protein matrix with the following concentrations of SDMA:

REF	Product Description	Quantity/Volume
5085-0003-00	**ARK SDMA Control** SDMA, buffer, bovine serum albumin, and preservatives (nominal level)	Dropper vials
	LOW (14.0 µg/dL)	1 X 2 mL
	HIGH (50.0 µg/dL)	1 X 2 mL

Each laboratory should establish its own ranges for each new lot of controls.

4 Warnings and Precautions

- For Veterinary Use only.
- Do not mix controls from different lot numbers.
- Use each lot as a set.
- Controls contain ≤0.09% sodium azide.

5 Instructions For Use

- For a complete summary and explanation of the ARK SDMA Assay, refer to the package insert for the ARK SDMA Assay, REF 5085-0001-00, REF 5085-0001-01 or REF 5085-0001-02.
- Controls are ready to use. Mix each level by gentle inversion before dispensing.
- Squeeze sufficient volume (~40µL/drop) into individual sample cups for each level. Consult instrument-specific sample volume requirements. Return caps to their original containers and keep tight.
- Store at 2-8°C. Use prior to the expiration date.

6 Limitations of Procedure

Accurate and reproducible results are dependent upon properly functioning instruments, reagents, calibrators, controls, storage of product as directed, and good laboratory technique.

7 Trademarks

ARK™ is a trademark of ARK Diagnostics, Inc.
Other brand or product names are trademarks of their respective holders.

ARK Diagnostics, Inc.
Fremont, CA 94538 USA

Revised October 2026
1600-1214-00 Rev 03